

CERTIFICATE OF ANALYSIS

Catalog No: ASTM-P-126-01

Description: Group 1,2 - Synthetic Distillation Standard

Lot: 221011398

Solvent: N/A

Hazards: Refer to SDS for complete safety information

Date Certified: Jan 27, 2021

Expiration: Jan 27, 2031

Sample Size: 500 mL

Storage Condition: Ambient (>5 °C)

Certified Reference Material



Signal Word: Danger



Volume	Degrees C		Degrees F	
	Average	STDS	Average	STDS
IBP	64.9	0.26	148.9	0.46
5%	82.2	0.15	179.9	0.21
10%	87.0	0.18	188.6	0.29
20%	93.2	0.18	199.8	0.26
30%	98.3	0.22	208.9	0.35
40%	103.6	0.26	218.5	0.42
50%	110.7	0.28	231.2	0.40
60%	118.8	0.45	245.8	0.76
70%	129.3	0.49	264.8	0.89
80%	142.6	0.72	288.7	1.35
90%	159.6	1.09	319.2	1.99
95%	181.7	2.03	359.2	3.69
Final	184.8	0.56	364.6	1.02

Distillation values are determined by ASTM D86.

A product with a suffix (-1A, -2B, etc. or -01, -02, etc.) on its lot number has had its expiration date extended and is identical to the same lot number without the suffix. Labels and certificates follow U.S. Conventions in reporting numerical values: A comma (,) is used to separate units of one-thousand or greater. A period (.) is used as a decimal place marker.

The information on this certificate may not be reproduced without the express permission of the manufacturer. See reverse side for additional information.

Hazard Information: Please refer to the SDS for information regarding the hazards associated with using this material.

This product was prepared according to in-house procedures and is guaranteed to be homogeneous.

Certified By:



Larry Decker, Organic QC Manager